



# Lab Supplies for PFAS Analysis

## Methods-Based Reference to Lab Supplies for PFAS Testing

PFAS “forever chemicals” are ubiquitous in the environment, but they are also commonly found in lab supplies and even in instruments, so it is essential to prevent background sources from contaminating samples. PFAS can also be “sticky” and adhere to certain materials that are regularly used in sampling and analysis (e.g., glass containers or pipettes), so everything in the sample flow path must be considered for its ability to either contribute background PFAS or to retain PFAS that is present in samples.

The risk of PFAS contamination or retention can make choosing the right lab supplies for PFAS analysis difficult, but it is also challenging because many methods are available, and they use different sample preparation and analysis products. Selecting suitable lab supplies is even harder when you want to measure trace levels of PFAS that are not listed in current government methods or guidelines, such as ultrashort- and short-chain PFAS.

Finding suitable products can be simplified using the product guide tables on the following pages, which detail the lab supplies for PFAS analysis that are appropriate for widely accepted test methods from around the world, including U.S. agencies (CDC, EPA, FDA, and USDA); European agencies (DIN); and international standards (ASTM and ISO). After selecting method-appropriate products, be sure to evaluate your system and supplies to ensure they are free of target analytes and qualified for PFAS analysis at the levels required by the method.

Restek is proud to support PFAS testing labs with clean, high-quality lab supplies for PFAS testing and expert technical support for the analysis of legacy, alternative, and ultrashort-chain PFAS across a wide range of global testing standards. Visit [www.restek.com/PFAS](http://www.restek.com/PFAS) to explore more resources for PFAS analysis. In addition, if you want to see results before setting foot in the lab, you can model PFAS methods in minutes from your desk using our free, online Pro EZLC chromatogram modeler.

## ANALYSIS

	EPA 537.1	EPA 533	EPA 1633	EPA OTM 45	EPA 8327	USDA CLG-PFAS 2.04	CDC 6304.09	US FDA C-010.03	DIN EN 17892	DIN 38407-42	ISO 21675:2019	ISO 23702-1:2023	ASTM D7968-23	ASTM D7979-20	ASTM D8421-22
<b>C18 PHASE COLUMNS</b>															
PFAS delay column: 5 µm, 50 x 2.1 mm (cat.# 27854)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
PFAS delay column: 5 µm, 50 x 3.0 mm (cat.# 580959 [custom product])	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Force C18 FPP column: 50 x 2.1 mm, 1.8 µm (cat.# 9634252)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Raptor C18 SPP column: 50 x 2.1 mm, 2.7 µm, (cat.# 9304A52)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
<i>More C18 column dimensions are available at <a href="http://www.restek.com">www.restek.com</a></i>															
<b>SPECIALTY PHASE COLUMNS</b>															
Raptor Polar X: 2.7 µm, 50 x 2.1 mm (cat.# 9311A52)	Ideal for ultrashort-chain (C1-C3) PFAS analysis with isocratic methods.														
Ultra IBD: 3 µm, 100 x 2.1 mm (cat.# 9175312)	Ideal for ultrashort-, short-, mid-, and long-chain PFAS analysis (C1 - C14) with gradient methods.														

**NOTE:** Some products shown above may not exactly match the product description in the methods.

## STANDARDS


For detailed compound lists, view these products on [www.restek.com](http://www.restek.com).

	EPA 537.1	EPA 533	EPA 1633	EPA OTM 45	EPA 8327	USDA CLG-PFAS 2.04	CDC 6304.09	US FDA C-010.03	DIN EN 17892	DIN 38407-42	ISO 21675:2019	ISO 23702-1:2023	ASTM D7968-23	ASTM D7979-20	ASTM D8421-22
PFAS 24 calibration standard: 1 µg/mL, methanol (1 mM KOH/2-propanol [98:2]) (cat.# 30733)	-	X	-	-	-	-	-	-	-	X	-	-	-	-	-
PFAS 28 calibration standard: 1 µg/mL, methanol (1 mM KOH/2-propanol [98:2]) (cat.# 30734)	-	X	-	-	-	-	-	-	-	X	-	-	-	-	-
EPA 537.1 PFAS calibration standard: 2 µg/mL, methanol (1 mM KOH) (cat.# 30735)	X	-	-	-	X	-	-	-	-	-	-	-	-	-	-
EPA 533 PFAS calibration standard: 2 µg/mL, methanol (1 mM KOH) (cat.# 30736)	-	X	-	-	X	-	-	-	-	X	-	-	-	-	-

**NOTE:** Some products shown above may not exactly match the product description in the methods.



Discover more solutions for PFAS analysis at [www.restek.com/PFAS!](http://www.restek.com/PFAS!)



## SAMPLE HANDLING

Product Category	Description	EPA 537.1	EPA 533	EPA 1633	EPA OTM 45	EPA 8327	USDA CLG-PFAS 2.04	CDC 6304.09	US FDA C-010.03	DIN EN 17892	DIN 38407-42	ISO 21675:2019	ISO 23702-1:2023	ASTM D7968-23	ASTM D7979-20	ASTM D8421-22
<b>PRE-SCREENING OR FILTRATION OF PARTICULATE-CONTAINING SAMPLES</b>																
Syringe filters	30 mm syringe filter: 0.45 µm, nylon (cat.# 23981)	-	-	X	-	X	-	-	X	-	X	X	-	-	-	-
	30 mm syringe filter: 0.22 µm, nylon (cat.# 23980)	-	-	X	-	X	-	-	X	-	X	X	-	-	-	-
	25 mm syringe filter: 0.22 µm, polypropylene (cat.# 28935)	-	-	-	-	-	-	-	-	-	-	-	X	X	X	X
Glass wool	Deactivated borosilicate glass wool (cat.# 20789)	-	-	X	X	-	-	-	-	-	-	-	-	-	-	-
Filter Vials	Thomson SINGLE StEP standard filter vials: 0.2 µm, nylon w/preslit cap (cat.# 25891)	-	-	X	-	-	-	-	-	-	-	-	-	-	-	-
	Thomson SINGLE StEP nano filter vials: 0.2 µm, nylon w/preslit cap (cat.# 25886) <sup>†</sup>	-	-	-	-	-	-	-	X	-	-	-	-	-	-	-
<i>More filter vial materials and formats are available at <a href="http://www.restek.com">www.restek.com</a></i>																
<b>SAMPLE COLLECTION, TRANSFER, AND EXTRACTION</b>																
SPE (WAX)	Resprep polymeric SPE cartridge: WAX, 6 mL/500 mg, 30 µm (cat.# 28291)	-	X	X	X	-	-	-	X	X	X	X	-	-	-	-
	Resprep polymeric SPE cartridge: WAX, 6 mL/200 mg, 30 µm (cat.# 28292)	-	X	X	X	-	-	-	X	X	X	X	-	-	-	-
<i>More WAX SPE configurations are available at <a href="http://www.restek.com">www.restek.com</a></i>																
SPE (S-DVB)	Resprep S-DVB SPE cartridge: 6 mL/500 mg (cat.# 28937)	X	-	-	-	-	-	*	-	X	-	-	-	-	-	-
Polypropylene sample reservoir	Polypropylene sample reservoir: 75 mL (cat.# 26015)	X	X	X	-	-	-	-	-	X	X	X	-	-	-	-
Polypropylene connectors	Polypropylene connectors (cat.# 26007)	X	X	X	-	-	-	-	-	X	X	X	-	-	-	-
QuEChERS extraction salts	Q-sep QuEChERS extraction salt packets only (original): 6 g MgSO <sub>4</sub> , 1.5 g NaCl (cat.# 582025)	-	-	-	-	-	-	-	X	-	-	-	-	-	-	-
<i>More QuEChERS products available at <a href="http://www.restek.com">www.restek.com</a></i>																
Centrifuge	Q-sep multispeed centrifuge for QuEChERS (cat.# 28295)	-	-	X	-	-	X	-	X	X	-	X	X	X	-	-
Ultraclean resin	Ultra-Clean resin (cat.# 24230)	-	-	-	X	-	-	-	-	-	-	-	-	-	-	-
Vacuum manifold	Resprep QR-24 vacuum manifold (cat.# 28299-VM)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
	Resprep QR-12 vacuum manifold (cat.# 28298-VM)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Vacuum pumps	Chemker 300 PTFE vacuum pump: 20 L/min, AC110 V, 60 Hz (cat.# 27426)	X	X	X	X	X	X	*	X	X	X	X	X	X	X	X
	Chemker 300 PTFE vacuum pump: 18 L/min, AC220 V, 50 Hz (cat.# 27427)	X	X	X	X	X	X	*	X	X	X	X	X	X	X	X

**NOTE:** Some products shown above may not exactly match the product description in the methods.

\* This CDC method is based on automated online SPE using C8 or C18 SPE materials. The product shown here is recommended for off-line SPE.

<sup>†</sup> This U.S. FDA method uses nano filter vial without caps. Cat.#. 25886 comes with preslit caps, but polyethylene caps (cat.#s 23244 and 23247) can be ordered separately.

X A regenerated cellulose filter is mentioned in the method. However, nylon and polypropylene are widely accepted filter membrane materials for PFAS analysis.

X A glass fiber filter is mentioned in the method. However, nylon and polypropylene are widely accepted filter membrane materials for PFAS analysis.

## GENERAL LAB SUPPLIES

	EPA 537.1	EPA 533	EPA 1633	EPA OTM-45	EPA 8327	USDA CLG-PFAS 2.04	CDC 6304.09	US FDA C-010.03	DIN EN 17892	DIN 38407-42	ISO 21675:2019	ISO 23702-1:2023	ASTM D7968-23	ASTM D7979-20	ASTM D8421-22
Polypropylene vials: 100-pk. (cat.# 23243)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Polypropylene vials: 1000-pk. (cat.# 23246)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Polyethylene vial caps: 100-pk. (cat.# 23244)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Polyethylene vial caps: 1000-pk. (cat.# 23247)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Centrifuge tubes: 50 mL, 50-pk. (cat.# 25846)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Centrifuge tubes: 50 mL, 500-pk. (cat.# 28290)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Norm-Ject plastic syringe: 10 mL Luer Lock tip (cat.# 22775)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
Norm-Ject plastic syringe: 5 mL Luer Lock tip (cat.# 22774)	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X

**NOTE:** Some products shown above may not exactly match the product description in the methods.

**X** ASTM methods refer to the use of a glass syringe. Norm-Ject plastic syringes are made from polypropylene/polyethylene and can be used as an alternative.

### Recommended Products for Volatile Fluorinated Compounds (VFCs) Analysis by U.S. EPA OTM-50 (GC-MS)

	EPA OTM-50
<b>SAMPLE COLLECTION/TRANSPORT</b>	
SilcoCan Air Sampling Canister, 6 L, with 3-Port Siltek-Treated RAVE+ Valve with Gauge (cat.# 27309)	X
SilcoCan Air Sampling Canister, 6 L, with 2-Port Siltek-Treated RAVE+ Valve (cat.# 27307)	X
SilcoCan Air Sampling Canister, 6 L, No Valve (cat.# 22092)	X
TO-Can air sampling canister, 6 L, with 3-port RAVE+ valve with gauge (cat.# 27319)	X
TO-Can air sampling canister, 6 L, with 2-port RAVE+ valve (cat.# 27318)	X
TO-Can air sampling canister, 6 L, no valve (cat.# 22096)	X

## References

- Restek Corporation, Guide to PFAS testing methods, a global survey of PFAS testing methods and guidelines, Technical article, EVAR3552C-UNV, 2024. <https://www.restek.com/articles/guide-to-pfas-testing-methods>
- S.H., Liang, J. A. Steimling, M. Chang Analysis of ultrashort-chain and short-chain (C1 to C4) per- and polyfluorinated substances in potable and non-potable waters. J. Chromatogr. Open 4 (100098) (2023) <https://doi.org/10.1016/j.jcoa.2023.100098>
- U.S. Environmental Protection Agency, EPA Method 537.1, Determination of selected per- and polyfluorinated alkyl substances in drinking water by solid phase extraction and liquid chromatography/tandem mass spectrometry (LC/MS/MS), November 2018. [https://cfpub.epa.gov/si/si\\_public\\_file\\_download.cfm?p\\_download\\_id=537290&Lab=NERL](https://cfpub.epa.gov/si/si_public_file_download.cfm?p_download_id=537290&Lab=NERL)
- U.S. Environmental Protection Agency, EPA Method 533, Determination of per- and polyfluoroalkyl substances in drinking water by isotope dilution anion exchange solid phase extraction and liquid chromatography/tandem mass spectrometry, November 2019. <https://www.epa.gov/sites/default/files/2019-12/documents/method-533-815b19020.pdf>
- U.S. Environmental Protection Agency, EPA Method 1633, Analysis of per- and polyfluoroalkyl substances (pfas) in aqueous, solid, biosolids, and tissue samples by LC-MS/MS, January 2024. <https://www.epa.gov/system/files/documents/2024-01/method-1633-final-for-web-posting.pdf>
- U.S. Environmental Protection Agency, Other Test Method-45 (OTM-45), Measurement of selected per- and polyfluorinated alkyl substances from stationary sources, January 13, 2021. [https://www.epa.gov/sites/default/files/2021-01/documents/otm\\_45\\_semi-volatile\\_pfas\\_1-13-21.pdf](https://www.epa.gov/sites/default/files/2021-01/documents/otm_45_semi-volatile_pfas_1-13-21.pdf)
- U.S. Environmental Protection Agency, EPA Method 8327, Per- and polyfluoroalkyl substances (PFAS) by liquid chromatography/tandem mass spectrometry (LC/MS/MS), July 2021. <https://www.epa.gov/system/files/documents/2021-07/8327.pdf>
- U.S. Department of Agriculture, Food Safety and Inspection Service, Chemistry Laboratory Guidebook, CLG-PFAS 2.04, Screening, determination, and confirmation of PFAS by UHPLC-MS-MS, February 28, 2023. [https://www.fsis.usda.gov/sites/default/files/media\\_file/documents/CLG-PFAS2.04.pdf](https://www.fsis.usda.gov/sites/default/files/media_file/documents/CLG-PFAS2.04.pdf)
- U.S. Centers for Disease Control and Prevention, CDC 6304.09, Perfluoroalkyl and polyfluoroalkyl substances, February 2019. <https://www.cdc.gov/nchs/data/nhanes/2017-2018/labmethods/PFAS-J-MET-508.pdf>
- U.S. Food and Drug Administration, FDA C-010.02, Determination of 30 per and polyfluoroalkyl substances (PFAS) in food and feed using liquid chromatography-tandem mass spectrometry (LC-MS/MS), April 12, 2024. <https://www.fda.gov/media/131510/download>
- DIN EN 17892, Water quality-Determination of the sum of perfluorinated substances (Sum of PFAS) in drinking water - Method using liquid chromatography/mass spectrometry (LC/MS), September 2021.
- DIN 38407-42, German standard methods for the examination of water, waste water and sludge-Jointly determinable substances (group F)-Part 42: determination of selected polyfluorinated compounds (PFC) in water-Method using high performance liquid chromatography and mass spectrometric detection (HPLC/MS-MS) after solid-liquid extraction (F 42), March 2011.
- ISO 21675:2019, Water quality—Determination of perfluoroalkyl and polyfluoroalkyl substances (PFAS) in water—Method using solid phase extraction and liquid chromatography-tandem mass spectrometry (LC-MS/MS), 2019. <https://www.iso.org/standard/71338.html>
- ASTM D7968-23, Standard test method for determination of polyfluorinated compounds in soil by liquid chromatography tandem mass spectrometry (LC/MS/MS), November 6, 2023. <https://www.astm.org/d7968-23.html>
- ASTM D7979-20, Standard test method for determination of per- and polyfluoroalkyl substances in water, sludge, influent, effluent, and wastewater by liquid chromatography tandem mass spectrometry (LC/MS/MS), September 3, 2020. <https://www.astm.org/d7979-20.html>
- ASTM D8421-22, Standard test method for determination of per- and polyfluoroalkyl substances (PFAS) in aqueous matrices by co-solvation followed by liquid chromatography tandem mass spectrometry (LC/MS/MS) <https://www.astm.org/d8421-22.html>
- U.S. Environmental Protection Agency, Other Test Method-50, OTM-50, Sampling and analysis of volatile fluorinated compounds from stationary sources using passivated stainless-steel canisters, January 2024. [https://www.epa.gov/system/files/documents/2024-01/otm-50-release-1\\_0.pdf](https://www.epa.gov/system/files/documents/2024-01/otm-50-release-1_0.pdf)

### Questions? Contact us or your local Restek representative ([www.restek.com/contact-us](http://www.restek.com/contact-us)).

Restek patents and trademarks are the property of Restek Corporation. (See [www.restek.com/Patents-Trademarks](http://www.restek.com/Patents-Trademarks) for full list.) Other trademarks in Restek literature or on its website are the property of their respective owners. Restek registered trademarks are registered in the U.S. and may also be registered in other countries. To unsubscribe from future Restek communications or to update your preferences, visit [www.restek.com/subscribe](http://www.restek.com/subscribe) to update your status with an authorized Restek distributor or instrument channel partner, please contact them directly.

© 2024 Restek Corporation. All rights reserved. Printed in the U.S.A.



[www.restek.com](http://www.restek.com)



Lit. Cat.# EVAR3498E-UNV